



Product Service

Attestation of Conformity

No. E8A 098238 0022 Rev. 00

Holder of Certificate: HANYOUNG NUX CO., LTD.

28, Gilpa-ro 71beon-gil, Nam-gu
Incheon 22121
REPUBLIC OF KOREA

Name of Object: Electrical equ. for measurement, control and laboratory use
(LCD Digital Temperature Controller)

Model(s): VX4-UCMA-abcdef, VX4-UCMA-A2D4, VX4-UCMA-A2D2H2, VX4-UCMA-A2CD2, VX4-UCMA-A2CH2, VX4-UCMA-A2CTR

Note: For the suffix information in the Above model name as follows;

- The "a" represents RELAY(sub output).
- The "b" represents RS-485(communication).
- The "c" represents DI(digital input).
- The "d" represents retransmission output.
- The "e" represents CT.
- The "f" represents remote input.

Description of Object:	Rated Input Voltage:	100-240 V a.c.
	Rated Frequency:	50/60 Hz
	Rated Power Consumption:	8.2 VA (Max)

Tested according to:	EN 61000-3-2:2014
	EN 61000-3-3:2013
	EN 61000-6-2:2005/AC:2005
	EN 61000-6-4:2007/A1:2011

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with all essential requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: DREKCEE1807-0503

Date, 2018-07-31

(Byung-Soo Kang)

Page 1 of 1

After preparation of the necessary technical documentation as well as the EU Declaration of conformity the required CE marking can be affixed on the product. That Declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.

TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD
ZERTIFIKAT ◆ CERTIFICATE ◆ 認證書 ◆ CERTIFICADO ◆ CERTIFICAT



Product Service

Attestation of Conformity

No. E8A 098238 0024 Rev. 00

Holder of Certificate: HANYOUNG NUX CO., LTD.

28, Gilpa-ro 71beon-gil, Nam-gu
Incheon 22121
REPUBLIC OF KOREA

Name of Object: Electrical equ. for measurement, control and laboratory use
(LCD Digital Temperature Controller)

Model(s): VX4-USMA-abcdef, VX4-USMA-A2D4, VX4-USMA-A2D2H2, VX4-USMA-A2CD2, VX4-USMA-A2CH2, VX4-USMA-A2CTR
Note: For the suffix information in the Above model name as follows;

- The "a" represents RELAY(sub output).
- The "b" represents RS-485(communication).
- The "c" represents DI(digital input).
- The "d" represents retransmission output.
- The "e" represents CT.
- The "f" represents remote input.

Description of Object:	Rated Input Voltage:	100-240 V a.c.
	Rated Frequency:	50/60 Hz
	Rated Power Consumption:	8.2 VA (Max)

Tested according to:	EN 61000-3-2:2014
	EN 61000-3-3:2013
	EN 61000-6-2:2005/AC:2005
	EN 61000-6-4:2007/A1:2011

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with all essential requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: DREKCEE1807-0505

Date, 2018-07-31 (Byung-Soo Kang)

Page 1 of 1

After preparation of the necessary technical documentation as well as the EU Declaration of conformity the required CE marking can be affixed on the product. That Declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.

ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ CERTIFICADO ◆ CERTIFICAT