



Attestation of Conformity

No. N8A 098238 0027 Rev. 00

Holder of Certificate: HANYOUNG NUX CO., LTD.

28, Gilpa-ro 71beon-gil, Nam-gu

Incheon 22121

REPUBLIC OF KOREA

Product: Electrical equ. for measurement, control and

laboratory use

(LCD Timer)

Model(s): LT4x

(x: Control output- S, blank)

Parameters: Rated input voltage: 24-240 V a.c. or 24-240 V d.c.

> 4.5 VA Rated input power:

> > 2 W

IPX₀

50/60 Hz Rated frequency:

Protection class: Class II

Degree of protection against ingress of liquids:

EN 61010-1:2010 **Tested**

according to:

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance.

It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: CPSA0131618

2018-09-13

(Harry Kwon)

Page 1 of 1

Date,

After preparation of the necessary technical documentation as well as the EU declaration of conformity the required CE marking can be affixed on the product. The declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.